

ALMEMO® 1020-2



Reference measuring instru-
ment for temperature
High-precision measuring
by means of thermocouples
Types N, S, R, B
Resolution 0.01 K, up to 1800 °C

Technical features

- Temperature measurement with very high levels of resolution, precision, and linearity, using thermocouples Types N, S, R, B
- Suitable as reference device in calibration laboratories and quality assurance procedures
- Very high accuracy thanks to multi-point adjustment of the thermocouple temperature sensor
- Each temperature sensor has its own cold junction stored in the ALMEMO® plug or externally. The cold junction temperature in the ALMEMO® plug is measured to a very high resolution of 0.001 K by means of an NTC sensor.
- Two electrically isolated measuring inputs for thermocouples, types N, S, R, B
- Resolution 0.01 K
- Units °C, °F, K
- High-resolution A/D converter, delta-sigma, 24-bit, 1.25 mops (measuring operations per second)
- Two output sockets for digital interface, ALMEMO® memory connector
- Compact, modern, ergonomic design
- Graphics display, illuminated with white light
- Easy and convenient to operate by means of 4 soft-keys and cursor block
- Measured value display : 2 measured values, differential, measuring point list, cold junction temperature
- Measuring functions : Zero-setting, smoothing, maximum / minimum values, individual value memory for 100 values
- Data logger with ALMEMO® memory connector (accessory)
- Sensor programming : Smoothing, designation, units
- Device configuration : Illumination, contrast, device address, baud rate
- Choice of language : German, English, French

Technical data ALMEMO® 1020-2

Measuring inputs	2 ALMEMO® input sockets for thermocouples	Outputs	2 ALMEMO® sockets for interface cable and ALMEMO® memory connector
Electrical isolation	Semiconductor relay (50 V)	Standard equipment	
A/D converter	Delta-sigma, 24-bit, 1.25 mops	Display	Graphics display, 128 x 64 pixels, 8 rows
Measuring ranges		Illumination	2 white LEDs
NiCrSi-NiSi Type N	-200 to +1300 °C	Keypad	7 silicone keys (of which 4 soft-keys)
PtRh10-Pt Type S	-50 to +1768 °C	Date and time-of-day	Real-time clock, buffered by battery
PtRh13-Pt Type R	-50 to +1768 °C	Individual value memory, internal	100 measured values
PtRh30-PtRh6 Type B	+250 to +1820 °C	Power supply	
Resolution	0.01 K	Battery set	3 AA alkaline batteries
Accuracy	±0.1 K ± 1 digit in range	Mains adapter	ZA1312NA10 100 to 240 VAC to 12 VDC, 2 A, electrically isolated
Type N	-200 to +1300 °C	Current consumption (without input and output modules)	approx. 20 mA
Type S	+50 to +1760 °C	With illumination	approx. 40 mA
Type R	+100 to +1760 °C	Housing	127 x 83 x 42 mm (LxWxH)
Type B	+500 to +1800 °C		ABS, 290g
Nominal conditions	23 °C ±2 K, 1013 mbar, battery mode		
Temperature drift	typical 10 ppm / K		
Cold junction temperature	Measuring operations with 0.001 K resolution		

Accessories	Order no.
Ethernet data cable	ZA1945DK
ALMEMO® memory connector with micro SD	ZA1904SD
Rubberized impact protection, gray	ZB2490GS2
DIN rail mounting	ZB2490HS

Variants

Complete set comprising reference measuring instrument for temperature plus accessories, evaluation software, thermocouple sensor, with DAkkS calibration certificate

Reference measuring instrument ALMEMO® 1020-2, including 3 AA alkaline batteries, mains unit ZA1312NA10, USB data cable ZA1919DKU, instrument case, and evaluation software ALMEMO® View SW5500AV (see page 06.16)

Set with high-precision sheathed thermocouple sensor type N



Set

with sheathed thermocouple sensor type N FTAN926L0500P2 with DAkkS calibration certificate at 0 / 100 / 500 / 1000 °C, including adjustment

Order no.

SP10202ND

Accessories: Aluminum profile case for 1 sensor (up to 500 mm in length)

ZB9000TK1

Technical data:

Sheathed thermocouple sensor type N FTAN926L0500P2

Measuring element NiCrSi-NiSi, type N, class 1

Measuring tip Mineral-insulated sheathed line, d = 6 mm, L = 500 mm

Operative range -200 to +1150 °C

Connecting cable 1.5 meters, thermal line (stranded wire) FEP / silicone (-50 to +200 °C)

ALMEMO® plug Resolution 0.01 K with integrated cold junction compensation sensor

Set with high-precision thermocouple sensor type S



Set

with thermocouple sensor type S FTAS916L0700P2 replacement ceramic protective tube, case for sensors ZB9000TK2 with DAkkS calibration certificate at 500/1000/1200 °C, including adjustment

Order no.

SP10202S1D

Technical data:

Thermocouple sensor type S FTAS916L0700P2

Measuring element PtRh10-Pt, Type S, Class 1

Measuring tip Thermowire, d = 0.5 mm in ceramic protective tube diameter = 6 mm, length = 700 mm up to +1400 °C

Operative range

Connection head ceramic protective tube, screwed

Connecting cable 1.5 meters, compensation line FEP / silicone (-50 to +200 °C)

ALMEMO® plug Resolution 0.01 K with integrated cold junction compensation sensor

Set with precision thermocouple sensor type S, with external cold junction



Set

with thermocouple sensor type S, with external cold junction FTAS906L0700P2, replacement ceramic protective tube, Case for sensors ZB9000TK2 with DAkkS calibration certificate at 500 / 1000 / 1200 °C, including adjustment

Order no.

SP10202S2D

Technical data:

Thermocouple sensor type S, with external cold junction FTAS906L0700P2

Measuring element PtRh10-Pt, Type S, Class 1

Measuring tip Thermowire, d = 0.5 mm in ceramic protective tube diameter = 6 mm, length = 700 mm up to +1600 °C

Operative range

Connection head ceramic protective tube, screwed

Connecting cable 0.75 meters, insulated, thermo-wires PtRh10-Pt as far as cold junction

Cold junction Stainless steel protective tube diameter = 5 mm, length = 250 mm

Connecting cable 2 meters, stranded copper wire

ALMEMO® plug Resolution 0.01 K

Certificates

06/2018 • We reserve the right to make technical changes.

Calibration certificate for ALMEMO® 1020-2 with precision sheathed thermocouple sensor type N
(Example)

the
gsstelle GmbH
bration laboratory in the
DKD
nst

Kalibrierzeichen
Calibration mark

Dieser Kalibrierschein dokumentiert die Durchführung der nationalen Kalibrierung auf der Grundlage der Darstellung der Einheiten in der SI-Einheit mit dem Internationalen Einheitsystem (SI). Die DAKKS ist Unterzeichner des Übereinkommens der Europäischen Union für die Anerkennung der gegenseitigen Kalibrierung (EA) und der Internationalen Laboratorien für die Anerkennung der gegenseitigen Kalibrierung (ILAC). Für die Einhaltung der Bestimmungen zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich. This calibration certificate is valid for the duration of the calibration.

entfühler, angeschlossen an
anzeigergerät
nnected with 1 temp. measuring device
ss- und Regelungstechnik GmbH

ALMEMO 1020-2
NiCrSi-NiSi, Typ N
H12070031

Kalibriergegenstand
1 Thermoelementfühler NiCrSi-NiSi, Typ N, Ø 6 mm Länge 760 mm, angeschlossen an ein Temperaturanzeigergerät ALMEMO 1020-2, Serien-Nr. H12070031

Object of calibration
1 thermocouple probe NiCrSi-NiSi, type N, Ø 6 mm length 760 mm, connected with one temperature measuring device ALMEMO 1020-2, Serial-No. H12070031

Messergebnisse / Test Result

Kanal Channel	Serien-Nr. Serial No.	Prüf­temperatur Test Temperature °C	Anzeige Indication °C	Abweichung Deviation K	Messunsicherheit Uncertainty K
M0	-	1150,00	1150,00	0,00	3,0
		1000,00	1000,00	0,00	1,5
		500,00	500,00	0,00	1,0
		100,00	100,00	0,00	0,3

Die Werte beziehen sich auf die Internationale Temperaturskala von 1990 (ITS-90).
The values are based on the International Temperature Scale of 1990 (ITS-90).

Calibration certificate for ALMEMO® 1020-2 with precision thermocouple sensor type S,
with external cold junction (Example)

the
gsstelle GmbH
bration laboratory in the
DKD
nst

Kalibrierzeichen
Calibration mark

Dieser Kalibrierschein dokumentiert die Durchführung der nationalen Kalibrierung auf der Grundlage der Darstellung der Einheiten in der SI-Einheit mit dem Internationalen Einheitsystem (SI). Die DAKKS ist Unterzeichner des Übereinkommens der Europäischen Union für die Anerkennung der gegenseitigen Kalibrierung (EA) und der Internationalen Laboratorien für die Anerkennung der gegenseitigen Kalibrierung (ILAC). Für die Einhaltung der Bestimmungen zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich. This calibration certificate is valid for the duration of the calibration.

entfühler mit externer
e, angeschlossen an
ranzeigergerät
le probe with external cold-junction,
th 1 temp. measuring device

less- und Regelungstechnik GmbH

ALMEMO 1020-2
Pt10%Rh-Pt, Typ S
H12070031

Kalibriergegenstand
1 Thermoelementfühler Pt10%Rh-Pt, Typ S, Schutzrohr: Keramik, Ø 8,2 mm, Länge 500 mm, mit externer Vergleichsstelle, angeschlossen an ein Temperaturanzeigergerät ALMEMO 1020-2, Serien-Nr. H12070031

Object of calibration
1 thermocouple probe Pt10%Rh-Pt, type S, Sheath tube: ceramics, Ø 8,2 mm, length 500 mm, with external cold-junction, connected with one temperature measuring device ALMEMO 1020-2, Serial-No. H12070031

Messergebnisse / Test Result

Kanal Channel	Serien-Nr. Serial No.	Prüf­temperatur Test Temperature °C	Anzeige Indication °C	Abweichung Deviation K	Messunsicherheit Uncertainty K
M0	12050001	1200,00	1200,00	0,00	1,5
		1000,00	1000,00	0,00	1,0
		500,00	500,00	0,00	0,5

Die Werte beziehen sich auf die Internationale Temperaturskala von 1990 (ITS-90).
The values are based on the International Temperature Scale of 1990 (ITS-90).

Die Korrektur der Messkette erfolgte über die Mehrpunktjustage-Funktion!
The correction of the measuring system was realized by the multiple point function!

Bedingungen während der Kalibrierung
Calibration Conditions

Other certificates for measuring instruments and sensors (see chapter „Calibration certificates“)